

Specification

Liquid version of the agar, recommended for studies of antibiotic MIC (Minimum Inhibitory Concentration).

Formula * in g/L

Peptone.....	17.50
Starch.....	1.50
Meat infusion solids.....	2.00

Final pH 7,3 ±0,1 at 25 °C

* Adjusted and /or supplemented as required to meet performance criteria

Directions

Add 21 g of powder to 1 L of distilled water and dissolve completely. Distribute in suitable containers. Sterilize in the autoclave at 121°C for 15 minutes.

Quality control

Incubation temperature: 37 °C ±1.0 **Incubation time:** 21± 3h

Inoculum: ≤100 CFU. Min. 50 CFU (Productivity)

Microorganism	Growth	Remarks
<i>Staphylococcus aureus</i> ATCC® 25923	Good	-
<i>Escherichia coli</i> ATCC® 25922	Good	-
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Good	-
<i>Escherichia coli</i> ATCC® 35218	Good	-
<i>Enterococcus faecalis</i> ATCC® 29212	Good	-

References

- BAUER, A.L., W.M.M. KIRBY, J.C. SHERRIS & M. TURCK (1966) Antibiotic susceptibility testing by a standardized single disc method. A. J. Clin. Pathol 45:493.
- BARRY, A.L., M.D. COYLE, C. THORNBERRY, E.H. GARLACH & R.W. HAWKINSON (1979) Methods of measuring zones of inhibition with Bauer-Kirby disk-susceptibility test. J. Clin. Microbiol. 10:885-889.
- CFR (1972) Rules and Regulations. 37:20525. Washington. DC. USA.
- CLSI (2006) Document M6-A2. Protocols for evaluating dehydrated Mueller-Hinton Agar: Approved Standard. 2nd ed. Clinical and Laboratory Standards Institute. Pennsylvania. USA.
- CLSI (2006) Document M2-A9. Performance standards for antimicrobial disk susceptibility tests: Approved Standard. 9th ed. Clinical and Laboratory Standards Institute. Pennsylvania. USA.
- ERICSSON & SHERRIS (1971) Acta Pathol. Microbiol. Scand Suppl 217 p:90.
- HINDLER, J. (1998) Antimicrobial Susceptibility Testing in Essential Procedures for Clinical Microbiology. ASM Press. Washington. DC. USA.
- MUNRO, S., (1995) Disk Difusión Susceptibility Testing, in Clinical Microbiology Procedures Handbook, H.D. Isenberg (ed) APHA. Washington. DC. USA.
- MILLER, J.M, C. THORNBERRY & C.N. BAKER (1984) Disk difusión susceptibility test troubleshooting guide. Lab. Med. 15:183-185.
- NEUMAN, M.A., D.F. SAMM, C. THORNBERRY & I.E. MCGOWAN (1991) New developments in antimicrobial agent susceptibility testing: A practical guide. ASM. Washington. DC. USA.

Storage

For laboratory use only. Keep tightly closed, away from bright light, in a cool dry place (+4 °C to 30 °C).